

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005923.D
 Acq On : 13 Nov 2018 02:12
 Operator : JC/MD
 Sample : J5918-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Nov 13 06:42:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	103728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	151131	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	133327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	63482	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	66365	55.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.30%	
35) Dibromofluoromethane	5.50	113	51248	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.72	98	175340	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.14	95	59374	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005923.D
Acq On : 13 Nov 2018 02:12
Operator : JC/MD
Sample : J5918-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Nov 13 06:42:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

