

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005927.D
 Acq On : 13 Nov 2018 03:59
 Operator : JC/MD
 Sample : J5892-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PN-STONE

Quant Time: Nov 13 06:52:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	101034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	146285	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132816	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	63156	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	63944	55.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.10%	
35) Dibromofluoromethane	5.50	113	50383	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
50) Toluene-d8	8.71	98	174695	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
62) 4-Bromofluorobenzene	11.14	95	59245	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
Target Compounds						
16) Acetone	2.45	43	3791	6.062	ug/l	93
20) Methylene Chloride	2.86	84	7893	7.683	ug/l	99
25) 2-Butanone	4.70	43	1950	1.983	ug/l	93
43) Isopropyl Acetate	6.46	43	71436	26.584	ug/l	99
95) Naphthalene	13.83	128	10959	3.533	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005927.D
Acq On : 13 Nov 2018 03:59
Operator : JC/MD
Sample : J5892-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PN-STONE

Quant Time: Nov 13 06:52:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

