

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005928.D
 Acq On : 13 Nov 2018 04:25
 Operator : JC/MD
 Sample : J5894-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DW-5

Quant Time: Nov 13 06:55:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	104059	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	148777	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132205	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	62447	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	63514	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
35) Dibromofluoromethane	5.50	113	51404	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
50) Toluene-d8	8.71	98	174710	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
62) 4-Bromofluorobenzene	11.14	95	58330	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
Target Compounds						
16) Acetone	2.45	43	14009	21.749	ug/l	97
18) Methyl Acetate	2.78	43	4419	1.110	ug/l	94
20) Methylene Chloride	2.85	84	7231	6.834	ug/l	97
25) 2-Butanone	4.70	43	6410	6.330	ug/l	97
43) Isopropyl Acetate	6.46	43	68782	25.168	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005928.D
Acq On : 13 Nov 2018 04:25
Operator : JC/MD
Sample : J5894-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DW-5

Quant Time: Nov 13 06:55:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

