

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005934.D
 Acq On : 13 Nov 2018 07:05
 Operator : JC/MD
 Sample : J5894-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DW-10

Quant Time: Nov 13 07:28:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	104028	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	160984	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	135181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	65300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	67360	56.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.64%	
35) Dibromofluoromethane	5.51	113	54023	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
50) Toluene-d8	8.71	98	179573	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.14	95	59439	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	8984	13.952	ug/l	96
18) Methyl Acetate	2.78	43	4427	1.115	ug/l	94
20) Methylene Chloride	2.85	84	5875	5.554	ug/l	97
25) 2-Butanone	4.70	43	4227	4.175	ug/l	92
43) Isopropyl Acetate	6.46	43	71032	24.020	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005934.D
Acq On : 13 Nov 2018 07:05
Operator : JC/MD
Sample : J5894-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DW-10

Quant Time: Nov 13 07:28:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

