

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111218\
 Data File : VX005937.D
 Acq On : 13 Nov 2018 08:25
 Operator : JC/MD
 Sample : J5907-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2-S

Quant Time: Nov 13 12:15:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	101948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	142205	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	131975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	61463	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	65533	55.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.82%	
35) Dibromofluoromethane	5.51	113	51180	52.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.84%	
50) Toluene-d8	8.71	98	175934	54.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.42%	
62) 4-Bromofluorobenzene	11.14	95	58587	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
Target Compounds						
16) Acetone	2.45	43	6836	10.833	ug/l	91
20) Methylene Chloride	2.85	84	13828	13.340	ug/l	93
43) Isopropyl Acetate	6.46	43	6766	2.590	ug/l	99
95) Naphthalene	13.83	128	2821	1.800	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111218\
Data File : VX005937.D
Acq On : 13 Nov 2018 08:25
Operator : JC/MD
Sample : J5907-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2-S

Quant Time: Nov 13 12:15:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

