

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013375.D
 Acq On : 12 Nov 2019 10:39
 Operator : JC/SP
 Sample : VX1112WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1112WBL01

Quant Time: Nov 12 12:49:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	132969	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	742082	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	646110	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.04	65	288805	29.41	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.03%
60) 4-Bromofluorobenzene	11.13	95	291159	28.04	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.47%
63) Toluene-d8	8.71	98	894904	30.63	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013375.D
Acq On : 12 Nov 2019 10:39
Operator : JC/SP
Sample : VX1112WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1112WBL01

Quant Time: Nov 12 12:49:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

