

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013383.D
 Acq On : 12 Nov 2019 14:22
 Operator : JC/SP
 Sample : VX1112WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1112WBL02

Quant Time: Nov 13 02:55:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	738307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1156293	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1020184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	454126	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	414549	47.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.78%	
35) Dibromofluoromethane	5.48	113	350131	48.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.56%	
50) Toluene-d8	8.71	98	1369338	50.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.13	95	452718	45.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013383.D
Acq On : 12 Nov 2019 14:22
Operator : JC/SP
Sample : VX1112WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1112WBL02

Quant Time: Nov 13 02:55:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

