

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013386.D
 Acq On : 12 Nov 2019 15:42
 Operator : JC/SP
 Sample : K5803-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS027-191111

Quant Time: Nov 13 03:02:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	756712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1184317	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1054105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	475319	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	434439	48.46	ug/l	0.00
Spiked Amount			Recovery	=	96.92%	
35) Dibromofluoromethane	5.49	113	359775	48.43	ug/l	0.00
Spiked Amount			Recovery	=	96.86%	
50) Toluene-d8	8.71	98	1413093	50.46	ug/l	0.00
Spiked Amount			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	11.13	95	468556	46.36	ug/l	0.00
Spiked Amount			Recovery	=	92.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	12555	2.808	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	138447	16.353	ug/l	91
44) Trichloroethene	7.21	130	27057	3.106	ug/l	93
64) Tetrachloroethene	9.33	164	17482	2.317	ug/l	88
65) Chlorobenzene	10.14	112	39287	1.841	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013386.D
Acq On : 12 Nov 2019 15:42
Operator : JC/SP
Sample : K5803-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS027-191111

Quant Time: Nov 13 03:02:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

