

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013388.D
 Acq On : 12 Nov 2019 16:28
 Operator : JC/SP
 Sample : K5803-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-191111

Quant Time: Nov 13 03:21:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	761425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1194800	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1068517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	485914	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	432668	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
35) Dibromofluoromethane	5.48	113	362142	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	8.71	98	1422147	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.13	95	481992	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
Target Compounds						
16) Acetone	2.43	43	31832	7.075	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	18260	2.143	ug/l	91
44) Trichloroethene	7.21	130	4553	0.518	ug/l	79
65) Chlorobenzene	10.14	112	5622	0.260	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013388.D
Acq On : 12 Nov 2019 16:28
Operator : JC/SP
Sample : K5803-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-191111

Quant Time: Nov 13 03:21:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

