

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013389.D
 Acq On : 12 Nov 2019 16:51
 Operator : JC/SP
 Sample : K5803-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-191111

Quant Time: Nov 13 03:25:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	781601	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1233329	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1084530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	495576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	447921	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
35) Dibromofluoromethane	5.48	113	369342	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
50) Toluene-d8	8.71	98	1458776	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.14	95	490782	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
Target Compounds						
16) Acetone	2.43	43	9820	2.126	ug/l #	87
27) cis-1,2-Dichloroethene	4.58	96	72231	8.260	ug/l	87
44) Trichloroethene	7.22	130	3388	0.373	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013389.D
Acq On : 12 Nov 2019 16:51
Operator : JC/SP
Sample : K5803-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-191111

Quant Time: Nov 13 03:25:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

