

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111219\
 Data File : VX013392.D
 Acq On : 12 Nov 2019 18:01
 Operator : JC/SP
 Sample : K5799-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LTM-TB08-SRI4

Quant Time: Nov 13 03:34:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	775029	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1226413	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1092874	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	508304	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	436259	47.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.02%	
35) Dibromofluoromethane	5.48	113	367191	47.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.48%	
50) Toluene-d8	8.71	98	1458904	50.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.62%	
62) 4-Bromofluorobenzene	11.14	95	496935	47.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.96%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111219\
Data File : VX013392.D
Acq On : 12 Nov 2019 18:01
Operator : JC/SP
Sample : K5799-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LTM-TB08-SRI4

Quant Time: Nov 13 03:34:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

