

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX111219\
 Data File : VX013374.D
 Acq On : 12 Nov 2019 09:58
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 12 12:36:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	107	0.00
2 M	Dichlorodifluoromethane	2.248	2.023	10.0	100	0.00
3 M	Chloromethane	2.450	2.286	6.7	104	0.00
4 M	Vinyl Chloride	2.784	2.481	10.9	100	0.00
5 M	Bromomethane	1.673	1.473	12.0	111	0.00
6 M	Chloroethane	1.530	1.408	8.0	106	0.00
7 M	Trichlorofluoromethane	2.997	2.536	15.4	98	0.00
8 T	Diethyl Ether	1.368	1.202	12.1	102	0.00
9	1,1,2-Trichlorotrifluoroeth	1.803	1.576	12.6	100	0.00
10 M	1,1-Dichloroethene	1.853	1.617	12.7	101	0.00
11	Methyl Iodide	2.288	1.692	26.0	90	0.00
12	Methyl Acetate	3.247	2.811	13.4	99	0.00
13 M	Acrolein	0.345	0.309	10.4	103	0.00
14 M	Acrylonitrile	1.194	1.007	15.7	97	0.00
15 M	Acetone	0.362	0.360	0.6	101	0.00
16 M	Carbon Disulfide	5.168	4.489	13.1	102	0.00
17	Allyl chloride	3.412	2.996	12.2	101	0.00
18 M	Methylene Chloride	2.102	1.839	12.5	103	0.00
19 M	trans-1,2-Dichloroethene	1.980	1.772	10.5	105	0.00
20 T	Diisopropyl ether	6.564	5.862	10.7	103	0.00
21 M	1,1-Dichloroethane	3.575	3.182	11.0	103	0.00
22 M	cis-1,2-Dichloroethene	2.258	2.003	11.3	104	0.00
23 M	tert-Butyl Alcohol	0.575	0.478	16.9	101	0.00
24 M	Methyl tert-Butyl Ether	6.132	5.540	9.7	104	0.00
25 M	Chloroform	3.500	3.144	10.2	104	0.00
26	Cyclohexane	3.247	2.767	14.8	100	0.00
27 s	1,2-Dichloroethane-d4	2.215	2.106	4.9	105	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
29	1,1-Dichloropropene	0.458	0.423	7.6	102	0.00
30 M	2-Butanone	0.302	0.293	3.0	100	0.00
31	2,2-Dichloropropane	0.496	0.477	3.8	103	0.00
32 M	1,1,1-Trichloroethane	0.509	0.473	7.1	101	0.00
33 M	Carbon Tetrachloride	0.442	0.399	9.7	97	0.00
34 M	Benzene	1.390	1.333	4.1	103	0.00
35	Methacrylonitrile	0.302	0.279	7.6	99	0.00
36 M	1,2-Dichloroethane	0.486	0.469	3.5	104	0.00
37 M	Trichloroethene	0.372	0.352	5.4	104	0.00
38	Methylcyclohexane	0.562	0.513	8.7	100	0.00
39 M	1,2-Dichloropropane	0.359	0.338	5.8	102	0.00
40	Dibromomethane	0.239	0.220	7.9	100	0.00
41 M	Bromodichloromethane	0.467	0.437	6.4	101	0.00
42 M	Vinyl Acetate	0.952	0.906	4.8	103	0.00
43	Ethyl Acetate	0.552	0.502	9.1	97	0.00
44	Isopropyl Acetate	0.887	0.816	8.0	101	0.00
45 T	1,4-Dioxane	0.010	0.009#	10.0	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX1111219\
 Data File : VX013374.D
 Acq On : 12 Nov 2019 09:58
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 12 12:36:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.441	0.408	7.5	103	0.00
47	n-amyl Acetate	0.778	0.720	7.5	104	0.00
48 M	t-1,3-Dichloropropene	0.524	0.489	6.7	106	0.00
49 T	cis-1,3-Dichloropropene	0.574	0.547	4.7	105	0.00
50 M	1,1,2-Trichloroethane	0.354	0.325	8.2	98	0.00
51	Ethyl methacrylate	0.583	0.538	7.7	104	0.00
52	1,3-Dichloropropane	0.605	0.573	5.3	101	0.00
53 M	Dibromochloromethane	0.375	0.351	6.4	104	0.00
54 M	1,2-Dibromoethane	0.378	0.357	5.6	101	0.00
55 M	2-Chloroethyl vinyl ether	0.249	0.231	7.2	99	0.00
56 M	Bromoform	0.298	0.269	9.7	104	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
58 M	4-Methyl-2-Pentanone	0.627	0.579	7.7	99	0.00
59 M	2-Hexanone	0.496	0.475	4.2	102	0.00
60 S	4-Bromofluorobenzene	0.482	0.486	-0.8	107	0.00
61 M	Tetrachloroethene	0.370	0.355	4.1	106	0.00
62 M	Toluene	1.671	1.590	4.8	102	0.00
63 S	Toluene-d8	1.357	1.380	-1.7	105	0.00
64 M	Chlorobenzene	1.045	0.999	4.4	106	0.00
65	1,1,1,2-Tetrachloroethane	0.391	0.364	6.9	103	0.00
66 M	Ethyl Benzene	1.836	1.736	5.4	103	0.00
67 M	m/p-Xylenes	0.708	0.670	5.4	104	0.00
68 M	o-Xylene	0.683	0.649	5.0	105	0.00
69 M	Styrene	1.167	1.098	5.9	104	0.00
70	Isopropylbenzene	1.801	1.714	4.8	104	0.00
71 M	1,1,2,2-Tetrachloroethane	0.645	0.602	6.7	102	0.00
72	1,2,3-Trichloropropane	0.544	0.568	-4.4	123	0.00
73	Bromobenzene	0.475	0.451	5.1	106	0.00
74	n-propylbenzene	2.047	1.947	4.9	105	0.00
75	2-Chlorotoluene	1.217	1.154	5.2	104	0.00
76	1,3,5-Trimethylbenzene	1.514	1.433	5.4	104	0.00
77	t-1,4-Dichloro-2-butene	0.220	0.195	11.4	103	0.00
78	4-Chlorotoluene	1.394	1.297	7.0	104	0.00
79	tert-butylbenzene	1.527	1.405	8.0	101	0.00
80	1,2,4-Trimethylbenzene	1.508	1.435	4.8	105	0.00
81	sec-Butylbenzene	1.742	1.645	5.6	105	0.00
82	p-Isopropyltoluene	1.605	1.525	5.0	106	0.00
83 M	1,3-Dichlorobenzene	0.835	0.799	4.3	109	0.00
84 M	1,4-Dichlorobenzene	0.829	0.784	5.4	108	0.00
85	n-Butylbenzene	1.386	1.286	7.2	106	0.00
86 T	Hexachloroethane	0.295	0.268	9.2	107	0.00
87 M	1,2-Dichlorobenzene	0.835	0.798	4.4	107	0.00
88	1,2-Dibromo-3-Chloropropane	0.144	0.127	11.8	102	0.00
89	1,2,4-Trichlorobenzene	0.576	0.536	6.9	107	0.00
90	Hexachlorobutadiene	0.276	0.251	9.1	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX1111219\
Data File : VX013374.D
Acq On : 12 Nov 2019 09:58
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Nov 12 12:36:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.823	1.696	7.0	105	0.00
92	1,2,3-Trichlorobenzene	0.582	0.538	7.6	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0