

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX111219\
 Data File : VX013395.D
 Acq On : 12 Nov 2019 19:11
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 13 02:48:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 M	Dichlorodifluoromethane	2.248	2.147	4.5	97	0.00
3 M	Chloromethane	2.450	2.384	2.7	100	0.00
4 M	Vinyl Chloride	2.784	2.678	3.8	99	0.00
5 M	Bromomethane	1.673	1.448	13.4	100	0.00
6 M	Chloroethane	1.530	1.465	4.2	101	0.00
7 M	Trichlorofluoromethane	2.997	2.796	6.7	99	0.00
8 T	Diethyl Ether	1.368	1.287	5.9	100	0.00
9	1,1,2-Trichlorotrifluoroeth	1.803	1.639	9.1	95	0.00
10 M	1,1-Dichloroethene	1.853	1.756	5.2	101	0.00
11	Methyl Iodide	2.288	2.015	11.9	99	0.00
12	Methyl Acetate	3.247	2.927	9.9	95	0.00
13 M	Acrolein	0.345	0.264	23.5	81	0.00
14 M	Acrylonitrile	1.194	1.035	13.3	92	0.00
15 M	Acetone	0.362	0.300	17.1	77	0.00
16 M	Carbon Disulfide	5.168	4.702	9.0	98	0.00
17	Allyl chloride	3.412	3.085	9.6	95	0.00
18 M	Methylene Chloride	2.102	2.002	4.8	103	0.00
19 M	trans-1,2-Dichloroethene	1.980	1.880	5.1	102	0.00
20 T	Diisopropyl ether	6.564	6.191	5.7	100	0.00
21 M	1,1-Dichloroethane	3.575	3.404	4.8	101	0.00
22 M	cis-1,2-Dichloroethene	2.258	2.149	4.8	102	0.00
23 M	tert-Butyl Alcohol	0.575	0.486	15.5	94	0.00
24 M	Methyl tert-Butyl Ether	6.132	5.752	6.2	99	0.00
25 M	Chloroform	3.500	3.358	4.1	102	0.00
26	Cyclohexane	3.247	2.933	9.7	97	0.00
27 s	1,2-Dichloroethane-d4	2.215	2.151	2.9	98	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
29	1,1-Dichloropropene	0.458	0.449	2.0	102	0.00
30 M	2-Butanone	0.302	0.273	9.6	88	0.00
31	2,2-Dichloropropane	0.496	0.464	6.5	95	0.00
32 M	1,1,1-Trichloroethane	0.509	0.500	1.8	100	0.00
33 M	Carbon Tetrachloride	0.442	0.427	3.4	98	0.00
34 M	Benzene	1.390	1.391	-0.1	101	0.00
35	Methacrylonitrile	0.302	0.286	5.3	96	0.00
36 M	1,2-Dichloroethane	0.486	0.484	0.4	101	0.00
37 M	Trichloroethene	0.372	0.366	1.6	102	0.00
38	Methylcyclohexane	0.562	0.541	3.7	99	0.00
39 M	1,2-Dichloropropane	0.359	0.355	1.1	101	0.00
40	Dibromomethane	0.239	0.235	1.7	100	0.00
41 M	Bromodichloromethane	0.467	0.444	4.9	96	0.00
42 M	Vinyl Acetate	0.952	0.920	3.4	98	0.00
43	Ethyl Acetate	0.552	0.500	9.4	91	0.00
44	Isopropyl Acetate	0.887	0.823	7.2	96	0.00
45 T	1,4-Dioxane	0.010	0.009#	10.0	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX1111219\
 Data File : VX013395.D
 Acq On : 12 Nov 2019 19:11
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Nov 13 02:48:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.441	0.411	6.8	98	0.00
47	n-amyl Acetate	0.778	0.711	8.6	97	0.00
48 M	t-1,3-Dichloropropene	0.524	0.480	8.4	98	0.00
49 T	cis-1,3-Dichloropropene	0.574	0.551	4.0	99	0.00
50 M	1,1,2-Trichloroethane	0.354	0.353	0.3	100	0.00
51	Ethyl methacrylate	0.583	0.552	5.3	100	0.00
52	1,3-Dichloropropane	0.605	0.592	2.1	99	0.00
53 M	Dibromochloromethane	0.375	0.356	5.1	100	0.00
54 M	1,2-Dibromoethane	0.378	0.370	2.1	99	0.00
55 M	2-Chloroethyl vinyl ether	0.249	0.219	12.0	88	0.00
56 M	Bromoform	0.298	0.271	9.1	99	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
58 M	4-Methyl-2-Pentanone	0.627	0.588	6.2	96	0.00
59 M	2-Hexanone	0.496	0.452	8.9	92	0.00
60 S	4-Bromofluorobenzene	0.482	0.473	1.9	98	0.00
61 M	Tetrachloroethene	0.370	0.377	-1.9	106	0.00
62 M	Toluene	1.671	1.650	1.3	100	0.00
63 S	Toluene-d8	1.357	1.371	-1.0	99	0.00
64 M	Chlorobenzene	1.045	1.030	1.4	103	0.00
65	1,1,1,2-Tetrachloroethane	0.391	0.369	5.6	99	0.00
66 M	Ethyl Benzene	1.836	1.787	2.7	101	0.00
67 M	m/p-Xylenes	0.708	0.688	2.8	101	0.00
68 M	o-Xylene	0.683	0.669	2.0	103	0.00
69 M	Styrene	1.167	1.132	3.0	101	0.00
70	Isopropylbenzene	1.801	1.781	1.1	102	0.00
71 M	1,1,2,2-Tetrachloroethane	0.645	0.620	3.9	99	0.00
72	1,2,3-Trichloropropane	0.544	0.499	8.3	102	0.00
73	Bromobenzene	0.475	0.461	2.9	102	0.00
74	n-propylbenzene	2.047	1.994	2.6	102	0.00
75	2-Chlorotoluene	1.217	1.182	2.9	101	0.00
76	1,3,5-Trimethylbenzene	1.514	1.456	3.8	100	0.00
77	t-1,4-Dichloro-2-butene	0.220	0.186	15.5	93	0.00
78	4-Chlorotoluene	1.394	1.331	4.5	101	0.00
79	tert-butylbenzene	1.527	1.456	4.6	99	0.00
80	1,2,4-Trimethylbenzene	1.508	1.461	3.1	102	0.00
81	sec-Butylbenzene	1.742	1.681	3.5	102	0.00
82	p-Isopropyltoluene	1.605	1.541	4.0	101	0.00
83 M	1,3-Dichlorobenzene	0.835	0.806	3.5	104	0.00
84 M	1,4-Dichlorobenzene	0.829	0.794	4.2	103	0.00
85	n-Butylbenzene	1.386	1.291	6.9	101	0.00
86 T	Hexachloroethane	0.295	0.265	10.2	100	0.00
87 M	1,2-Dichlorobenzene	0.835	0.807	3.4	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.144	0.127	11.8	97	0.00
89	1,2,4-Trichlorobenzene	0.576	0.526	8.7	100	0.00
90	Hexachlorobutadiene	0.276	0.254	8.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX1111219\
Data File : VX013395.D
Acq On : 12 Nov 2019 19:11
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Nov 13 02:48:46 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.823	1.692	7.2	100	0.00
92	1,2,3-Trichlorobenzene	0.582	0.539	7.4	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0