

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111221\
 Data File : VX025155.D
 Acq On : 12 Nov 2021 15:51
 Operator : JC/MD
 Sample : M4615-08 10X
 Misc : 2.58g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COV14

Quant Time: Nov 13 04:24:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 12 12:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	213972	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	195873	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	104574	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	62022	42.905	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.820%
7) Chloroethane-d5	1.666	69	37288	45.260	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.520%
11) 1,1-Dichloroethene-d2	2.300	63	81043	32.597	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.200%
21) 2-Butanone-d5	4.465	46	99623	91.197	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.200%
24) Chloroform-d	5.056	84	113378	44.584	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.160%
26) 1,2-Dichloroethane-d4	5.958	65	70158	45.515	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.020%
32) Benzene-d6	5.970	84	237486	44.423	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.840%
36) 1,2-Dichloropropane-d6	7.305	67	73724	45.165	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.320%
41) Toluene-d8	8.647	98	223996	43.876	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.760%
43) trans-1,3-Dichloroprop...	8.951	79	39494	44.570	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.140%
47) 2-Hexanone-d5	9.384	63	78104	89.104	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.100%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	101127	43.013	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.020%
66) 1,2-Dichlorobenzene-d4	12.323	152	92721	44.728	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.460%
Target Compounds						
					Qvalue	
13) Acetone	2.392	43	33939	33.121	ug/L	99
22) 2-Butanone	4.568	43	3155	2.313	ug/L #	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111221\
Data File : VX025155.D
Acq On : 12 Nov 2021 15:51
Operator : JC/MD
Sample : M4615-08 10X
Misc : 2.58g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0V14

Quant Time: Nov 13 04:24:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 12 12:01:23 2021
Response via : Initial Calibration

