

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111221\
 Data File : VX025149.D
 Acq On : 12 Nov 2021 13:12
 Operator : JC/MD
 Sample : M4645-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Nov 13 04:23:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 12 12:01:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	203925	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	186959	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	98022	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	63303	45.949	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.900%
7) Chloroethane-d5	1.672	69	44588	56.787	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.580%
11) 1,1-Dichloroethene-d2	2.306	63	82102	34.650	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.300%
21) 2-Butanone-d5	4.464	46	107674	103.423	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.420%
24) Chloroform-d	5.056	84	115371	47.602	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	5.958	65	73954	50.341	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.680%
32) Benzene-d6	5.976	84	243259	47.673	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.340%
36) 1,2-Dichloropropane-d6	7.311	67	76308	48.976	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.960%
41) Toluene-d8	8.653	98	230060	47.212	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.420%
43) trans-1,3-Dichloroprop...	8.951	79	40940	48.405	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.800%
47) 2-Hexanone-d5	9.384	63	82934	99.125	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.130%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	107711	47.998	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	93044	47.884	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111221\
Data File : VX025149.D
Acq On : 12 Nov 2021 13:12
Operator : JC/MD
Sample : M4645-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Nov 13 04:23:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 12 12:01:23 2021
Response via : Initial Calibration

