

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111224\
 Data File : VX043809.D
 Acq On : 12 Nov 2024 14:45
 Operator : JC/MD
 Sample : VIBLK678
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK678

Quant Time: Nov 13 02:06:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 13 02:04:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	214273	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	192898	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	85251	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	45200	39.597	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.200%
7) Chloroethane-d5	1.642	69	29021	59.146	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	118.300%
11) 1,1-Dichloroethene-d2	2.294	65	22818	43.301	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.600%
21) 2-Butanone-d5	4.459	46	90106	120.558	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.560%
24) Chloroform-d	5.050	84	124412	50.557	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.120%
26) 1,2-Dichloroethane-d4	5.952	65	87123	55.731	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.460%
32) Benzene-d6	5.964	84	245897	49.306	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.620%
36) 1,2-Dichloropropane-d6	7.299	67	78019	51.420	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.840%
41) Toluene-d8	8.647	98	216714	47.835	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.680%
43) trans-1,3-Dichloroprop...	8.946	79	33378	45.999	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.000%
47) 2-Hexanone-d5	9.384	63	71725	119.448	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	119.450%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	120429	61.861	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	123.720%#
66) 1,2-Dichlorobenzene-d4	12.317	152	88254	59.749	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111224\
Data File : VX043809.D
Acq On : 12 Nov 2024 14:45
Operator : JC/MD
Sample : VIBLK678
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK678

Quant Time: Nov 13 02:06:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 13 02:04:29 2024
Response via : Initial Calibration

