

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111225\
 Data File : VX048572.D
 Acq On : 12 Nov 2025 12:40
 Operator : JC/MD
 Sample : Q3604-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-1

Quant Time: Nov 12 12:58:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	178690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	357430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	368888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	192403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	134160	53.882	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 107.760%		
35) Dibromofluoromethane	5.373	113	119142	44.041	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 88.080%		
50) Toluene-d8	8.635	98	411680	48.793	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 97.580%		
62) 4-Bromofluorobenzene	11.067	95	172063	54.722	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 109.440%		
Target Compounds						Qvalue
16) Acetone	2.374	43	100102	84.933	ug/l	95
20) Methylene Chloride	2.788	84	17291	7.136	ug/l	87
25) 2-Butanone	4.556	43	9928	5.694	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111225\
Data File : VX048572.D
Acq On : 12 Nov 2025 12:40
Operator : JC/MD
Sample : Q3604-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-1

Quant Time: Nov 12 12:58:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

