

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111225\
 Data File : VX048573.D
 Acq On : 12 Nov 2025 13:00
 Operator : JC/MD
 Sample : Q3604-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-2

Quant Time: Nov 12 13:18:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	212518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	429611	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	445317	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	214536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	168989	57.066	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.140%
35) Dibromofluoromethane	5.373	113	147905	45.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.980%
50) Toluene-d8	8.634	98	521695	51.444	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.880%
62) 4-Bromofluorobenzene	11.061	95	209428	55.415	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.840%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	99526	71.002	ug/l	99
20) Methylene Chloride	2.794	84	18055	6.265	ug/l	98
25) 2-Butanone	4.556	43	10024	4.834	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111225\
Data File : VX048573.D
Acq On : 12 Nov 2025 13:00
Operator : JC/MD
Sample : Q3604-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-2

Quant Time: Nov 12 13:18:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

