

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005957.D
 Acq On : 13 Nov 2018 19:00
 Operator : JC/MD
 Sample : J5923-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MS-OILY-STONE

Quant Time: Nov 14 03:28:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	91528	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	135849	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	125652	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	60260	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	63214	60.08	ug/l	0.00
Spiked Amount			Recovery	=	120.16%	
35) Dibromofluoromethane	5.50	113	49942	54.05	ug/l	0.00
Spiked Amount			Recovery	=	108.10%	
50) Toluene-d8	8.71	98	163834	53.33	ug/l	0.00
Spiked Amount			Recovery	=	106.66%	
62) 4-Bromofluorobenzene	11.14	95	54958	49.76	ug/l	0.00
Spiked Amount			Recovery	=	99.52%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	298	0.275	ug/l	90
11) Tert butyl alcohol	3.01	59	864	2.866	ug/l #	72
16) Acetone	2.45	43	5233	9.237	ug/l	100
18) Methyl Acetate	2.78	43	3944	1.149	ug/l	99
20) Methylene Chloride	2.86	84	11213	12.049	ug/l	88
25) 2-Butanone	4.70	43	851	0.955	ug/l	95
29) Tetrahydrofuran	5.16	42	401	0.701	ug/l #	50
31) Cyclohexane	5.67	56	1773	1.210	ug/l #	1
36) 1,1-Dichloropropene	5.67	75	6536	4.991	ug/l #	50
37) Ethyl Acetate	4.70	43	851	0.494	ug/l #	69
38) Carbon Tetrachloride	5.66	117	8517	6.496	ug/l #	15
41) Methacrylonitrile	5.17	41	210	0.226	ug/l #	42
43) Isopropyl Acetate	6.46	43	57598	23.081	ug/l	98
48) Methyl methacrylate	7.68	41	2402	2.144	ug/l #	24
51) 4-Methyl-2-Pentanone	8.62	43	1346	0.856	ug/l #	38
54) cis-1,3-Dichloropropene	8.62	75	346	0.262	ug/l #	71
59) 2-Hexanone	9.54	43	3520	2.539	ug/l #	55
74) N-amyl acetate	10.91	43	19	1.260	ug/l #	44
76) 1,2,3-Trichloropropane	11.14	75	28035	22.848	ug/l #	36
80) 1,3,5-Trimethylbenzene	11.51	105	712	0.240	ug/l	97
81) trans-1,4-Dichloro-2-buten	11.14	75	28035	72.820	ug/l #	9
95) Naphthalene	13.83	128	2718	1.790	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005957.D
Acq On : 13 Nov 2018 19:00
Operator : JC/MD
Sample : J5923-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MS-OILY-STONE

Quant Time: Nov 14 03:28:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :


MS-OILY-STONE

Tot Ion:168 Resp: 91528

Ion Ratio Lower Upper

168 100

99 43.7 40.7 61.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

40000

30000



20000

10000

0

Time-->

5.60 5.70 5.80



#3

Chloromethane

Concen: 0.275 ug/l

RT: 1.33 min Scan# 29

Delta R.T. 0.01 min

Lab File: VX005957.D


Acq: 13 Nov 2018 19:00

Tgt Ion: 50 Resp: 298

Ion Ratio Lower Upper

50 100

52 38.1 26.2 39.2



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.33

250

200

150


100


50

0

Time-->

1.30 1.35









#11

Tert butyl alcohol
 Concen: 2.866 ug/l
 RT: 3.01 min Scan# 305
 Delta R.T. -0.06 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Instrument :
 MSVOA_X
 ClientSampleId :
 MS-OILY-STONE




Tot Ion: 59 Resp: 864
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.5 12.7#




#16

Acetone
 Concen: 9.237 ug/l
 RT: 2.45 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Tgt Ion: 43 Resp: 5233
 Ion Ratio Lower Upper
 43 100
 58 29.8 23.8 35.6









#18
Methyl Acetate
Concen: 1.149 ug/l
RT: 2.78 min Scan# 266
Delta R.T. -0.01 min
Lab File: VX005957.D
Acq: 13 Nov 2018 19:00




Instrument :
MSVOA_X
ClientSampleId :
MS-OILY-STONE


Tot Ion: 43 Resp: 3944
Ion Ratio Lower Upper
43 100
74 20.7 16.8 25.2



#20
Methylene Chloride
Concen: 12.049 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX005957.D
Acq: 13 Nov 2018 19:00

Tgt Ion: 84 Resp: 11213
Ion Ratio Lower Upper
84 100
49 159.0 109.5 164.3
51 42.0 33.3 49.9
86 69.6 52.7 79.1





#25

2-Butanone

Concen: 0.955 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :


MS-OILY-STONE

Tot Ion: 43 Resp: 851

Ion Ratio Lower Upper


43 100

72 26.3 19.2 28.8





Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0



Time-->



#29

Tetrahydrofuran

Concen: 0.701 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.01 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00


Tgt Ion: 42 Resp: 401

Ion Ratio Lower Upper

42 100

72 0.0 32.0 48.0#


71 17.2 29.8 44.8#




Abundance Ion 42.00 (41.70 to 42.70): VX0


Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



Time-->





#31

Cyclohexane

Concen: 1.210 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.09 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion: 56 Resp: 1773

Ion Ratio Lower Upper

56 100

69 177.1 22.6 34.0#




84 119.3 62.7 94.1#



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 84.00 (83.70 to 84.70): VX0



#33

1,2-Dichloroethane-d4

Concen: 60.076 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX005957.D


Acq: 13 Nov 2018 19:00

Tgt Ion: 65 Resp: 63214

Ion Ratio Lower Upper



65 100


67 54.9 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Instrument :
 MSVOA_X
 ClientSampleId :
 MS-OILY-STONE

Tot Ion:114 Resp: 135849
 Ion Ratio Lower Upper
 114 100
 63 20.3 0.0 42.8
 88 15.5 0.0 31.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 54.051 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.01 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Tgt Ion:113 Resp: 49942
 Ion Ratio Lower Upper
 113 100
 111 98.9 82.3 123.5
 192 22.5 17.9 26.9



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.991 ug/l
 RT: 5.67 min Scan# 740
 Delta R.T. -0.13 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Instrument :
 MSVOA_X
 ClientSampleId :
 MS-OILY-STONE

Tot Ion: 75 Resp: 6536
 Ion Ratio Lower Upper
 75 100
 110 9.4 17.4 52.2#
 77 0.0 25.8 38.6#




Abundance Ion 74.90 (74.60 to 75.60): VX0
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VX0






#37
 Ethyl Acetate
 Concen: 0.494 ug/l
 RT: 4.70 min Scan# 582
 Delta R.T. -0.15 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Tgt Ion: 43 Resp: 851
 Ion Ratio Lower Upper
 43 100
 61 0.0 10.8 16.2#
 70 0.0 7.8 11.8#



Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 70.00 (69.70 to 70.70): VX0





#38

Carbon Tetrachloride

Concen: 6.496 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.12 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion:117 Resp: 8517

Ion Ratio Lower Upper

117 100

119 5.4 78.1 117.1#




121 0.0 24.6 36.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.226 ug/l

RT: 5.17 min Scan# 658

Delta R.T. 0.12 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Tgt Ion: 41 Resp: 210


Ion Ratio Lower Upper

41 100

39 33.3 42.2 63.2#

67 0.0 53.4 80.0#

52 0.0 23.4 35.2#






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0





#43

Isopropyl Acetate

Concen: 23.081 ug/l

RT: 6.46 min Scan# 871

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion: 43 Resp: 57598

Ion Ratio Lower Upper




43 100

61 19.2 14.6 22.0

87 12.4 9.6 14.4



Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 61.00 (60.70 to 61.70): VX0
Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 2.144 ug/l

RT: 7.68 min Scan# 1071

Delta R.T. -0.09 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00


Tgt Ion: 41 Resp: 2402

Ion Ratio Lower Upper



41 100


69 0.0 63.2 94.8#

39 14.7 42.2 63.2#



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0





#50

Toluene-d8

Concen: 53.334 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :


MS-OILY-STONE

Tot Ion: 98 Resp: 163834

Ion Ratio Lower Upper




98 100

100 63.6 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



#51

4-Methyl-2-Pentanone

Concen: 0.856 ug/l

RT: 8.62 min Scan# 1225

Delta R.T. -0.03 min

Lab File: VX005957.D


Acq: 13 Nov 2018 19:00

Tgt Ion: 43 Resp: 1346

Ion Ratio Lower Upper



43 100


58 0.0 29.7 44.5#



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#54
 cis-1,3-Dichloropropene
 Concen: 0.262 ug/l
 RT: 8.62 min Scan# 1224
 Delta R.T. 0.18 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00




Instrument :
 MSVOA_X
 ClientSampleId :
 MS-OILY-STONE


Tot Ion: 75 Resp: 346
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.2 37.8#
 39 40.2 38.2 57.2



#59
 2-Hexanone
 Concen: 2.539 ug/l
 RT: 9.54 min Scan# 1376
 Delta R.T. 0.04 min
 Lab File: VX005957.D
 Acq: 13 Nov 2018 19:00

Tgt Ion: 43 Resp: 3520
 Ion Ratio Lower Upper
 43 100
 58 20.0 25.9 77.8#





#62

4-Bromofluorobenzene

Concen: 49.761 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion: 95 Resp: 54958

Ion Ratio Lower Upper

95 100

174 93.0 0.0 184.4




176 88.1 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VX005825.D (-1631) (-)

Ion 173.80 (173.50 to 174.50): VX005825.D (-1631) (-)

Ion 175.80 (175.50 to 176.50): VX005825.D (-1631) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.12 min Scan# 1470

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00


Tgt Ion: 117 Resp: 125652

Ion Ratio Lower Upper

117 100

82 55.3 44.1 66.1



119 32.1 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): VX005825.D (-1462) (-)

Ion 82.00 (81.70 to 82.70): VX005825.D (-1462) (-)

Ion 118.90 (118.60 to 119.60): VX005825.D (-1462) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion:152 Resp: 60260

Ion Ratio Lower Upper

152 100

115 55.7 39.4 118.1




150 154.7 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 1.260 ug/l

RT: 10.91 min Scan# 1600

Delta R.T. 0.01 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Tgt Ion: 43 Resp: 19


Ion Ratio Lower Upper

43 100

70 0.0 31.6 47.4#

55 0.0 19.7 29.5#

61 0.0 18.3 27.5#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#76

1,2,3-Trichloropropane

Concen: 22.848 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.16 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :


MS-OILY-STONE

Tot Ion: 75 Resp: 28035

Ion Ratio Lower Upper

75 100

77 1.3 20.5 61.6#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0

11.14

25000



20000

15000

10000

5000

Time-->



#80

1,3,5-Trimethylbenzene

Concen: 0.240 ug/l

RT: 11.51 min Scan# 1698

Delta R.T. -0.00 min

Lab File: VX005957.D


Acq: 13 Nov 2018 19:00

Tgt Ion:105 Resp: 712

Ion Ratio Lower Upper

105 100

120 46.2 24.1 72.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

11.51

600

500

400


300


200

100

0

Time-->





#81

trans-1,4-Dichloro-2-butene

Concen: 72.820 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.07 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00

Instrument :

MSVOA_X

ClientSampleId :

MS-OILY-STONE


Tot Ion: 75 Resp: 28035

Ion Ratio Lower Upper




75 100

53 0.1 80.2 120.2#

89 0.0 39.8 59.8#



Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 53.00 (52.70 to 53.70): VX0
Ion 88.90 (88.60 to 89.60): VX0



#95

Naphthalene

Concen: 1.790 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. 0.00 min

Lab File: VX005957.D

Acq: 13 Nov 2018 19:00


Tgt Ion: 128 Resp: 2718

Ion Ratio Lower Upper

128 100

127 13.4 10.4 15.6

129 9.6 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

