

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005960.D
 Acq On : 13 Nov 2018 20:20
 Operator : JC/MD
 Sample : J5931-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS036-181112

Quant Time: Nov 14 04:39:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	91122	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	130868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	118782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	57018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	59479	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
35) Dibromofluoromethane	5.49	113	47780	53.68	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	107.36%	
50) Toluene-d8	8.71	98	156314	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
62) 4-Bromofluorobenzene	11.14	95	52082	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
Target Compounds						
16) Acetone	2.45	43	7948	14.091	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	5560	5.927	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005960.D
Acq On : 13 Nov 2018 20:20
Operator : JC/MD
Sample : J5931-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS036-181112

Quant Time: Nov 14 04:39:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

