

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005961.D
 Acq On : 13 Nov 2018 20:47
 Operator : JC/MD
 Sample : J5931-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-181112

Quant Time: Nov 14 04:41:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	97919	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	6.87	114	138239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	124980	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	60451	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	65436	58.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.26%	
35) Dibromofluoromethane	5.51	113	51530	54.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.62%	
50) Toluene-d8	8.71	98	166909	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	
62) 4-Bromofluorobenzene	11.14	95	55588	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
Target Compounds						
16) Acetone	2.44	43	3386	5.586	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005961.D
Acq On : 13 Nov 2018 20:47
Operator : JC/MD
Sample : J5931-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-181112

Quant Time: Nov 14 04:41:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

