

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005962.D
 Acq On : 13 Nov 2018 21:14
 Operator : JC/MD
 Sample : J5913-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DOC-19-PRE-CHAR-COMPOSITE

Quant Time: Nov 14 04:44:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	88768	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	130650	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	119846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	56187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	60600	59.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.76%	
35) Dibromofluoromethane	5.50	113	46329	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
50) Toluene-d8	8.71	98	156865	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
62) 4-Bromofluorobenzene	11.14	95	51820	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
Target Compounds						
16) Acetone	2.45	43	5969	10.863	ug/l	99
20) Methylene Chloride	2.86	84	33265	36.855	ug/l	89
43) Isopropyl Acetate	6.46	43	58722	24.468	ug/l	99
95) Naphthalene	13.83	128	3256	1.964	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005962.D
Acq On : 13 Nov 2018 21:14
Operator : JC/MD
Sample : J5913-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DOC-19-PRE-CHAR-COMPOSITE

Quant Time: Nov 14 04:44:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

