

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005963.D
 Acq On : 13 Nov 2018 21:40
 Operator : JC/MD
 Sample : J5934-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-0142

Quant Time: Nov 14 04:47:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	95520	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	132801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	123032	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	60550	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	62611	57.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.04%	
35) Dibromofluoromethane	5.51	113	48470	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
50) Toluene-d8	8.71	98	161703	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
62) 4-Bromofluorobenzene	11.14	95	55770	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	24914	21.957	ug/l	100
21) trans-1,2-Dichloroethene	3.17	96	8582	9.637	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	14668	14.917	ug/l	88
40) Benzene	6.15	78	4781	1.281	ug/l	95
65) Chlorobenzene	10.14	112	47652	19.334	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005963.D
Acq On : 13 Nov 2018 21:40
Operator : JC/MD
Sample : J5934-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-0142

Quant Time: Nov 14 04:47:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

