

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111318\
 Data File : VX005964.D
 Acq On : 13 Nov 2018 22:07
 Operator : JC/MD
 Sample : J5934-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-TB-2018-4

Quant Time: Nov 14 09:31:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	96480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	134091	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	122953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	59832	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	63696	57.43	ug/l	0.00
Spiked Amount			Recovery	=	114.86%	
35) Dibromofluoromethane	5.51	113	47076	51.62	ug/l	0.00
Spiked Amount			Recovery	=	103.24%	
50) Toluene-d8	8.72	98	162016	53.43	ug/l	0.00
Spiked Amount			Recovery	=	106.86%	
62) 4-Bromofluorobenzene	11.14	95	54709	50.19	ug/l	0.00
Spiked Amount			Recovery	=	100.38%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111318\
Data File : VX005964.D
Acq On : 13 Nov 2018 22:07
Operator : JC/MD
Sample : J5934-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-TB-2018-4

Quant Time: Nov 14 09:31:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

