

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111319\
 Data File : VX013399.D
 Acq On : 13 Nov 2019 11:22
 Operator : JC/SP
 Sample : VX1113WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1113WBL01

Quant Time: Nov 14 04:42:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	745716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1178063	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1030773	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	469469	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	425619	48.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.36%	
35) Dibromofluoromethane	5.48	113	354132	47.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.86%	
50) Toluene-d8	8.71	98	1397393	50.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.34%	
62) 4-Bromofluorobenzene	11.13	95	461900	45.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111319\
Data File : VX013399.D
Acq On : 13 Nov 2019 11:22
Operator : JC/SP
Sample : VX1113WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1113WBL01

Quant Time: Nov 14 04:42:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

