

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111319\
 Data File : VX013402.D
 Acq On : 13 Nov 2019 13:11
 Operator : JC/SP
 Sample : K5802-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0674

Quant Time: Nov 14 07:15:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	787325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1254405	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1100157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	482735	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	449940	48.24	ug/l	0.00
Spiked Amount						
			Recovery	=	96.48%	
35) Dibromofluoromethane	5.48	113	378829	48.15	ug/l	0.00
Spiked Amount						
			Recovery	=	96.30%	
50) Toluene-d8	8.71	98	1484766	50.06	ug/l	0.00
Spiked Amount						
			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	11.13	95	487155	45.51	ug/l	0.00
Spiked Amount						
			Recovery	=	91.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.19	85	9873	1.088	ug/l	90
4) Vinyl Chloride	1.40	62	395076	35.276	ug/l	99
12) 1,1-Dichloroethene	2.36	96	11180	1.548	ug/l	92
16) Acetone	2.42	43	9695	2.084	ug/l	92
17) Carbon Disulfide	2.56	76	8731	0.452	ug/l	98
21) trans-1,2-Dichloroethene	3.15	96	39393	5.010	ug/l	97
27) cis-1,2-Dichloroethene	4.58	96	571960	64.931	ug/l	89
44) Trichloroethene	7.20	130	61372	6.651	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111319\
Data File : VX013402.D
Acq On : 13 Nov 2019 13:11
Operator : JC/SP
Sample : K5802-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
S32-0674

Quant Time: Nov 14 07:15:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

