

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111319\
 Data File : VX013404.D
 Acq On : 13 Nov 2019 13:58
 Operator : JC/SP
 Sample : K5802-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0675

Quant Time: Nov 14 07:18:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	745728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1191329	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1058323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	489804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	425952	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	5.49	113	359025	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
50) Toluene-d8	8.71	98	1405310	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
62) 4-Bromofluorobenzene	11.14	95	475640	46.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.58%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.19	85	8273	0.963	ug/l	84
4) Vinyl Chloride	1.40	62	19367	1.826	ug/l	99
16) Acetone	2.42	43	10054	2.282	ug/l #	87
17) Carbon Disulfide	2.56	76	5673	0.310	ug/l #	91
21) trans-1,2-Dichloroethene	3.15	96	21166	2.842	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	246355	29.527	ug/l	89
44) Trichloroethene	7.21	130	74740	8.529	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111319\
Data File : VX013404.D
Acq On : 13 Nov 2019 13:58
Operator : JC/SP
Sample : K5802-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0675

Quant Time: Nov 14 07:18:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

