

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111320\
 Data File : VX019424.D
 Acq On : 13 Nov 2020 13:20
 Operator : JC/MD
 Sample : VX1113WBL01
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK301

Quant Time: Nov 17 04:29:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 04:26:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	17338	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.09	117	15079	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.06	152	6906	0.50	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.39	65	9374	0.51	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.00%
4) 1,2-Dichloroethane-d4	6.03	65	7185	0.44	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	88.00%
7) 1,2-Dichloropropane-d6	7.37	67	7606	0.55	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.00%
8) Toluene-d8	8.68	98	18465	0.48	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.00%
10) 1,1,2,2-Tetrachloroethane-	11.22	84	5440	0.56	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	112.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111320\
Data File : VX019424.D
Acq On : 13 Nov 2020 13:20
Operator : JC/MD
Sample : VX1113WBL01
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK301

Quant Time: Nov 17 04:29:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 04:26:19 2020
Response via : Initial Calibration

