

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111320\
 Data File : VX019418.D
 Acq On : 13 Nov 2020 10:04
 Operator : JC/MD
 Sample : VSTD0.505
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5305

Quant Time: Nov 17 04:15:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 04:13:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	18135	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.09	117	15969	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.06	152	7664	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.39	65	9835	0.55	ug/L	0.00
4) 1,2-Dichloroethane-d4	6.03	65	7820	0.42	ug/L	0.00
7) 1,2-Dichloropropane-d6	7.37	67	7912	0.58	ug/L	0.00
8) Toluene-d8	8.68	98	20010	0.40	ug/L	0.00
10) 1,1,2,2-Tetrachloroethane-	11.22	84	5779	0.62	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Vinyl chloride	1.39	62	9028	0.533	ug/L	100
6) Trichloroethene	7.18	95	6962	0.562	ug/L	100
9) 1,2-Dibromoethane	9.66	107	4082	0.562	ug/L	100
12) 1,2,3-Trichloropropane	11.27	75	3935	0.589	ug/L	100
13) 1,2-Dibromo-3-chloropropan	12.97	75	674	0.579	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111320\
Data File : VX019418.D
Acq On : 13 Nov 2020 10:04
Operator : JC/MD
Sample : VSTD0.505
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5305

Quant Time: Nov 17 04:15:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 04:13:39 2020
Response via : Initial Calibration

