

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038774.D
 Acq On : 13 Nov 2023 12:27
 Operator : JC/MD
 Sample : 05241-20DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EVAL-01D-20231102DL

Quant Time: Nov 14 05:31:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	290389	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	475032	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	499485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	289976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	162789	45.029	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.060%
35) Dibromofluoromethane	5.385	113	153369	44.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.660%
50) Toluene-d8	8.647	98	572744	46.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.560%
62) 4-Bromofluorobenzene	11.079	95	247737	52.191	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.380%
Target Compounds						
44) Trichloroethene	7.123	130	151138	39.682	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
Data File : VX038774.D
Acq On : 13 Nov 2023 12:27
Operator : JC/MD
Sample : 05241-20DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EVAL-01D-20231102DL

Quant Time: Nov 14 05:31:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

