

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038775.D
 Acq On : 13 Nov 2023 12:50
 Operator : JC/MD
 Sample : 05357-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20231107

Quant Time: Nov 14 05:31:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	282415	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	470811	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	495072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	287895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	162792	46.301	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.600%
35) Dibromofluoromethane	5.385	113	153828	45.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	8.647	98	563004	46.399	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	246844	52.469	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.940%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	13265	8.958	ug/l	94
20) Methylene Chloride	2.788	84	4097	1.373	ug/l #	76
25) 2-Butanone	4.574	43	3615	1.685	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
Data File : VX038775.D
Acq On : 13 Nov 2023 12:50
Operator : JC/MD
Sample : 05357-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20231107

Quant Time: Nov 14 05:31:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

