

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038776.D
 Acq On : 13 Nov 2023 13:13
 Operator : JC/MD
 Sample : 05358-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20231108DL

Quant Time: Nov 14 05:31:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	281923	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	462977	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	490894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	291534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	160777	45.807	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.620%		
35) Dibromofluoromethane	5.385	113	151634	45.474	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.940%		
50) Toluene-d8	8.647	98	554534	46.474	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.940%		
62) 4-Bromofluorobenzene	11.079	95	243924	52.726	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.460%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	6004	2.186	ug/l	Qvalue 93
12) 1,1-Dichloroethene	2.325	96	1229	0.467	ug/l	# 78
20) Methylene Chloride	2.794	84	1614	0.542	ug/l	# 75
27) cis-1,2-Dichloroethene	4.501	96	1240	0.375	ug/l	92
44) Trichloroethene	7.123	130	133406	35.939	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
Data File : VX038776.D
Acq On : 13 Nov 2023 13:13
Operator : JC/MD
Sample : 05358-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20231108DL

Quant Time: Nov 14 05:31:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

