

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038783.D
 Acq On : 13 Nov 2023 16:30
 Operator : JC/MD
 Sample : 05350-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT149S1-20231107RE

Quant Time: Nov 14 05:32:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168653	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	173636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	61560	59.522	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	= 119.040%#	
35) Dibromofluoromethane	5.385	113	52882	43.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 87.080%	
50) Toluene-d8	8.647	98	207718	47.788	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.580%	
62) 4-Bromofluorobenzene	11.079	95	96961	57.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 115.080%	
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.111	73	36165	14.587	ug/l	96
44) Trichloroethene	7.122	130	5803	4.291	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
Data File : VX038783.D
Acq On : 13 Nov 2023 16:30
Operator : JC/MD
Sample : 05350-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT149S1-20231107RE

Quant Time: Nov 14 05:32:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

