

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038769.D
 Acq On : 13 Nov 2023 10:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 05:29:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.519	0.517	0.4	96	0.00
3 P	Chloromethane	0.449	0.400	10.9	93	0.00
4 C	Vinyl Chloride	0.478	0.464	2.9#	97	0.00
5 T	Bromomethane	0.369	0.357	3.3	93	0.00
6 T	Chloroethane	0.303	0.285	5.9	96	0.00
7 T	Trichlorofluoromethane	0.824	0.846	-2.7	102	0.00
8 T	Diethyl Ether	0.283	0.269	4.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.487	0.496	-1.8	103	0.00
10 T	Methyl Iodide	0.601	0.661	-10.0	105	0.00
11 T	Tert butyl alcohol	0.127	0.128	-0.8	109	0.00
12 CM	1,1-Dichloroethene	0.467	0.457	2.1#	99	0.00
13 T	Acrolein	0.096	0.103	-7.3	113	0.00
14 T	Allyl chloride	0.668	0.630	5.7	96	0.00
15 T	Acrylonitrile	0.269	0.258	4.1	96	0.00
16 T	Acetone	0.262	0.328	-25.2#	133	0.00
17 T	Carbon Disulfide	1.240	1.140	8.1	92	0.00
18 T	Methyl Acetate	1.117	1.061	5.0	95	0.00
19 T	Methyl tert-butyl Ether	1.492	1.480	0.8	99	0.00
20 T	Methylene Chloride	0.528	0.498	5.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.494	2.8	98	0.00
22 T	Diisopropyl ether	1.260	1.230	2.4	97	-0.01
23 T	Vinyl Acetate	0.894	0.895	-0.1	97	0.00
24 P	1,1-Dichloroethane	0.833	0.805	3.4	97	0.00
25 T	2-Butanone	0.380	0.390	-2.6	105	-0.01
26 T	2,2-Dichloropropane	0.728	0.748	-2.7	105	0.00
27 T	cis-1,2-Dichloroethene	0.587	0.578	1.5	99	0.00
28 T	Bromochloromethane	0.375	0.366	2.4	99	0.00
29 T	Tetrahydrofuran	0.236	0.224	5.1	95	-0.01
30 C	Chloroform	0.978	0.931	4.8#	99	-0.01
31 T	Cyclohexane	0.689	0.664	3.6	96	-0.01
32 T	1,1,1-Trichloroethane	0.861	0.861	0.0	100	0.00
33 S	1,2-Dichloroethane-d4	0.622	0.570	8.4	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.360	0.349	3.1	98	0.00
36 T	1,1-Dichloropropene	0.447	0.437	2.2	98	0.00
37 T	Ethyl Acetate	0.457	0.437	4.4	94	-0.01
38 T	Carbon Tetrachloride	0.500	0.516	-3.2	101	0.00
39 T	Methylcyclohexane	0.559	0.556	0.5	100	0.00
40 TM	Benzene	1.288	1.253	2.7	97	0.00
41 T	Methacrylonitrile	0.233	0.225	3.4	95	0.00
42 TM	1,2-Dichloroethane	0.465	0.465	0.0	98	0.00
43 T	Isopropyl Acetate	0.702	0.688	2.0	96	-0.01
44 TM	Trichloroethene	0.401	0.403	-0.5	100	0.00
45 C	1,2-Dichloropropane	0.311	0.311	0.0	98	0.00
46 T	Dibromomethane	0.260	0.256	1.5	96	0.00
47 T	Bromodichloromethane	0.490	0.488	0.4	100	0.00
48 T	Methyl methacrylate	0.338	0.338	0.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
 Data File : VX038769.D
 Acq On : 13 Nov 2023 10:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 14 05:29:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 08 03:19:07 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	96	0.00
50 S	Toluene-d8	1.289	1.233	4.3	97	0.00
51 T	4-Methyl-2-Pentanone	0.457	0.455	0.4	97	0.00
52 CM	Toluene	0.844	0.860	-1.9#	99	0.00
53 T	t-1,3-Dichloropropene	0.499	0.528	-5.8	101	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.549	-4.4	100	0.00
55 T	1,1,2-Trichloroethane	0.358	0.366	-2.2	99	0.00
56 T	Ethyl methacrylate	0.501	0.536	-7.0	101	0.00
57 T	1,3-Dichloropropane	0.558	0.562	-0.7	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.246	0.247	-0.4	94	0.00
59 T	2-Hexanone	0.370	0.377	-1.9	100	0.00
60 T	Dibromochloromethane	0.402	0.426	-6.0	103	0.00
61 T	1,2-Dibromoethane	0.402	0.410	-2.0	100	0.00
62 S	4-Bromofluorobenzene	0.500	0.503	-0.6	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.374	0.370	1.1	101	0.00
65 PM	Chlorobenzene	1.018	1.047	-2.8	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.383	0.401	-4.7	103	0.00
67 C	Ethyl Benzene	1.722	1.768	-2.7#	101	0.00
68 T	m/p-Xylenes	0.683	0.703	-2.9	100	0.00
69 T	o-Xylene	0.665	0.682	-2.6	100	0.00
70 T	Styrene	1.097	1.154	-5.2	102	0.00
71 P	Bromoform	0.338	0.358	-5.9	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.067	3.129	-2.0	102	0.00
74 T	N-amyl acetate	1.066	1.071	-0.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.060	1.040	1.9	100	0.00
76 T	1,2,3-Trichloropropane	0.990	0.894	9.7	91	0.00
77 T	Bromobenzene	0.842	0.847	-0.6	103	0.00
78 T	n-propylbenzene	3.570	3.628	-1.6	101	0.00
79 T	2-Chlorotoluene	2.137	2.151	-0.7	102	0.00
80 T	1,3,5-Trimethylbenzene	2.640	2.740	-3.8	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.318	0.321	-0.9	101	0.00
82 T	4-Chlorotoluene	2.497	2.564	-2.7	102	0.00
83 T	tert-Butylbenzene	2.675	2.774	-3.7	102	0.00
84 T	1,2,4-Trimethylbenzene	2.618	2.716	-3.7	102	0.00
85 T	sec-Butylbenzene	3.338	3.386	-1.4	101	0.00
86 T	p-Isopropyltoluene	2.844	2.996	-5.3	104	0.00
87 T	1,3-Dichlorobenzene	1.610	1.629	-1.2	103	0.00
88 T	1,4-Dichlorobenzene	1.654	1.658	-0.2	103	0.00
89 T	n-Butylbenzene	2.530	2.593	-2.5	102	0.00
90 T	Hexachloroethane	0.482	0.517	-7.3	105	0.00
91 T	1,2-Dichlorobenzene	1.572	1.574	-0.1	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.267	0.268	-0.4	100	0.00
93 T	1,2,4-Trichlorobenzene	1.124	1.152	-2.5	104	0.00
94 T	Hexachlorobutadiene	0.485	0.500	-3.1	108	0.00
95 T	Naphthalene	3.391	3.475	-2.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111323\
Data File : VX038769.D
Acq On : 13 Nov 2023 10:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 14 05:29:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110723W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 08 03:19:07 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.108	1.107	0.1	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5