

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
 Data File : VX043834.D
 Acq On : 13 Nov 2024 15:29
 Operator : JC/MD
 Sample : P4785-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 14 04:30:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 13 02:04:29 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	269071	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	234504	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	104005	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	77857	54.315	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	= 108.620%		
7) Chloroethane-d5	1.642	69	48674	78.997	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	= 158.000%#		
11) 1,1-Dichloroethene-d2	2.294	65	36325	54.894	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	= 109.780%		
21) 2-Butanone-d5	4.458	46	82116	87.493	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	= 87.490%		
24) Chloroform-d	5.050	84	154253	49.918	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 99.840%		
26) 1,2-Dichloroethane-d4	5.952	65	99357	50.613	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 101.220%		
32) Benzene-d6	5.970	84	308888	50.947	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	= 101.900%		
36) 1,2-Dichloropropane-d6	7.305	67	88430	47.941	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	= 95.880%		
41) Toluene-d8	8.647	98	283277	51.434	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	= 102.860%		
43) trans-1,3-Dichloroprop...	8.951	79	40533	45.949	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	= 91.900%		
47) 2-Hexanone-d5	9.384	63	67403	92.335	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	= 92.330%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	118183	49.936	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	= 99.880%		
66) 1,2-Dichlorobenzene-d4	12.317	152	103559	57.469	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	= 114.940%		
Target Compounds						
13) Acetone	2.380	43	3962	4.539	ug/L	Qvalue 93
20) cis-1,2-Dichloroethene	4.483	96	3108	1.620	ug/L	93
61) 1,2,3-Trichloropropane	12.018	75	10774	5.337	ug/L #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
Data File : VX043834.D
Acq On : 13 Nov 2024 15:29
Operator : JC/MD
Sample : P4785-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 14 04:30:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 13 02:04:29 2024
Response via : Initial Calibration

