

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
 Data File : VX043835.D
 Acq On : 13 Nov 2024 15:53
 Operator : JC/MD
 Sample : P4785-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 14 04:30:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 13 02:04:29 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.757	114	244916	50.000	ug/L	0.00
28) Chlorobenzene-d5		10.055	117	211543	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		12.018	152	92869	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.368	65	76164	58.374	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	116.740%	
7) Chloroethane-d5		1.642	69	49013	87.392	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	174.780%#	
11) 1,1-Dichloroethene-d2		2.294	65	34852	57.863	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	115.720%	
21) 2-Butanone-d5		4.465	46	81473	95.369	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.370%	
24) Chloroform-d		5.050	84	151476	53.854	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.700%	
26) 1,2-Dichloroethane-d4		5.952	65	99039	55.427	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.860%	
32) Benzene-d6		5.964	84	309236	56.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.080%	
36) 1,2-Dichloropropane-d6		7.306	67	87212	52.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.820%	
41) Toluene-d8		8.647	98	280587	56.475	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.960%	
43) trans-1,3-Dichloroprop...		8.952	79	40657	51.092	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.180%	
47) 2-Hexanone-d5		9.385	63	63525	96.468	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.470%	
56) 1,1,2,2-Tetrachloroeth...		11.189	84	117380	54.980	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.960%	
66) 1,2-Dichlorobenzene-d4		12.317	152	100768	62.625	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	125.260%#	
Target Compounds							
61) 1,2,3-Trichloropropane		12.018	75	9301	5.160	ug/L	Qvalue # 60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
Data File : VX043835.D
Acq On : 13 Nov 2024 15:53
Operator : JC/MD
Sample : P4785-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 14 04:30:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM110424WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 13 02:04:29 2024
Response via : Initial Calibration

