

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
 Data File : VX043837.D
 Acq On : 13 Nov 2024 16:39
 Operator : JC/MD
 Sample : P4785-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 14 04:30:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 13 02:04:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	260968	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	229190	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	98622	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	72400	52.076	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 104.160%			
7) Chloroethane-d5	1.642	69	47195	78.975	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 157.940%#			
11) 1,1-Dichloroethene-d2	2.294	65	32837	51.164	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 102.320%			
21) 2-Butanone-d5	4.458	46	78854	86.626	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 86.630%			
24) Chloroform-d	5.050	84	147657	49.267	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 98.540%			
26) 1,2-Dichloroethane-d4	5.952	65	98914	51.952	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 103.900%			
32) Benzene-d6	5.970	84	297227	50.161	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.320%			
36) 1,2-Dichloropropane-d6	7.305	67	84071	46.635	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 93.260%			
41) Toluene-d8	8.647	98	270562	50.264	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 100.520%			
43) trans-1,3-Dichloroprop...	8.951	79	38788	44.991	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 89.980%			
47) 2-Hexanone-d5	9.384	63	63785	89.404	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 89.400%			
56) 1,1,2,2-Tetrachloroeth...	11.189	84	115514	49.940	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 99.880%			
66) 1,2-Dichlorobenzene-d4	12.317	152	97587	57.111	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 114.220%			
Target Compounds						Qvalue
61) 1,2,3-Trichloropropane	12.018	75	10894	5.691	ug/L	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111324\
Data File : VX043837.D
Acq On : 13 Nov 2024 16:39
Operator : JC/MD
Sample : P4785-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Nov 14 04:30:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML110424WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 13 02:04:29 2024
Response via : Initial Calibration

