

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
 Data File : VX048592.D
 Acq On : 13 Nov 2025 14:01
 Operator : JC/MD
 Sample : Q3609-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-713-COMP-01

Quant Time: Nov 14 03:57:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	154788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	326089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	350421	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	178918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	139972	64.897	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 129.800%#		
35) Dibromofluoromethane	5.373	113	114141	46.248	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.500%		
50) Toluene-d8	8.635	98	406462	52.805	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 105.600%		
62) 4-Bromofluorobenzene	11.061	95	167998	58.565	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 117.120%		
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	18649	18.266	ug/l	100
18) Methyl Acetate	2.709	43	3227	1.150	ug/l	92
20) Methylene Chloride	2.788	84	10219	4.869	ug/l	87
43) Isopropyl Acetate	6.330	43	12791	1.956	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
Data File : VX048592.D
Acq On : 13 Nov 2025 14:01
Operator : JC/MD
Sample : Q3609-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-713-COMP-01

Quant Time: Nov 14 03:57:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

