

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
 Data File : VX048593.D
 Acq On : 13 Nov 2025 14:22
 Operator : JC/MD
 Sample : Q3609-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-713-COMP-02

Quant Time: Nov 14 03:58:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	143432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	298042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	322250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	165942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	129948	65.019	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 130.040%	#
35) Dibromofluoromethane	5.373	113	105043	46.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.140%	
50) Toluene-d8	8.635	98	375623	53.391	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.780%	
62) 4-Bromofluorobenzene	11.061	95	177910	67.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 135.720%	#
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	15395	16.273	ug/l	99
18) Methyl Acetate	2.703	43	3121	1.200	ug/l #	90
20) Methylene Chloride	2.788	84	10041	5.162	ug/l #	86
43) Isopropyl Acetate	6.330	43	14339	2.399	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
Data File : VX048593.D
Acq On : 13 Nov 2025 14:22
Operator : JC/MD
Sample : Q3609-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-713-COMP-02

Quant Time: Nov 14 03:58:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

