

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
 Data File : VX048596.D
 Acq On : 13 Nov 2025 15:25
 Operator : JC/MD
 Sample : Q3609-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-713-COMP-05

Quant Time: Nov 14 03:59:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	145802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	297128	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	325195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	167787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	128418	63.209	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 126.420%#		
35) Dibromofluoromethane	5.373	113	101312	45.051	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 90.100%		
50) Toluene-d8	8.635	98	361763	51.579	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.160%		
62) 4-Bromofluorobenzene	11.061	95	154470	59.098	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 118.200%		
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	19146	19.909	ug/l	98
18) Methyl Acetate	2.703	43	3674	1.390	ug/l	97
20) Methylene Chloride	2.788	84	13606	6.882	ug/l	97
43) Isopropyl Acetate	6.330	43	12981	2.179	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
Data File : VX048596.D
Acq On : 13 Nov 2025 15:25
Operator : JC/MD
Sample : Q3609-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-713-COMP-05

Quant Time: Nov 14 03:59:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

