

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
 Data File : VX048598.D
 Acq On : 13 Nov 2025 16:06
 Operator : JC/MD
 Sample : Q3609-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-713-COMP-02

Quant Time: Nov 14 04:25:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	149491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	306571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	328681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	173162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	127530	61.223	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 122.440%#	
35) Dibromofluoromethane	5.379	113	103008	44.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 88.780%	
50) Toluene-d8	8.635	98	358415	49.527	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.060%	
62) 4-Bromofluorobenzene	11.061	95	153667	56.980	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	14761	14.970	ug/l	89
18) Methyl Acetate	2.703	43	3429	1.265	ug/l #	88
20) Methylene Chloride	2.788	84	10322	5.092	ug/l #	88
43) Isopropyl Acetate	6.330	43	13715	2.231	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
Data File : VX048598.D
Acq On : 13 Nov 2025 16:06
Operator : JC/MD
Sample : Q3609-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-713-COMP-02

Quant Time: Nov 14 04:25:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

