

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
 Data File : VX048599.D
 Acq On : 13 Nov 2025 16:27
 Operator : JC/MD
 Sample : Q3609-15RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-713-COMP-05RE

Quant Time: Nov 14 04:25:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	151779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	312476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	417156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	176357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	125800	59.482	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 118.960%#		
35) Dibromofluoromethane	5.373	113	104235	44.074	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 88.140%		
50) Toluene-d8	8.635	98	442067	59.932	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 119.860%#		
62) 4-Bromofluorobenzene	11.067	95	151921	55.268	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 110.540%		
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	15641	15.624	ug/l	97
18) Methyl Acetate	2.703	43	3281	1.193	ug/l	97
20) Methylene Chloride	2.788	84	10443	5.074	ug/l	86
43) Isopropyl Acetate	6.324	43	13853	2.211	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111325\
Data File : VX048599.D
Acq On : 13 Nov 2025 16:27
Operator : JC/MD
Sample : Q3609-15RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-713-COMP-05RE

Quant Time: Nov 14 04:25:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

