

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005972.D
 Acq On : 14 Nov 2018 14:22
 Operator : JC/MD
 Sample : J5925-05 10X
 Misc : 17.71g/5ml/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1D1

Quant Time: Nov 15 00:45:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	95865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	139642	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	125743	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	66981	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	60978	55.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.66%	
35) Dibromofluoromethane	5.50	113	49765	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
50) Toluene-d8	8.71	98	162955	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
62) 4-Bromofluorobenzene	11.14	95	59712	52.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.20%	
Target Compounds						
73) Isopropylbenzene	11.02	105	2161	0.56	ug/l	98
78) n-propylbenzene	11.36	91	4655	1.02	ug/l	99
85) sec-Butylbenzene	11.94	105	7429	1.82	ug/l	87
89) n-Butylbenzene	12.39	91	6185	1.85	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005972.D
Acq On : 14 Nov 2018 14:22
Operator : JC/MD
Sample : J5925-05 10X
Misc : 17.71g/5ml/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1D1

Quant Time: Nov 15 00:45:00 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

