

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005978.D
 Acq On : 14 Nov 2018 17:02
 Operator : JC/MD
 Sample : PB114787ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114787ZHE#01

Quant Time: Nov 15 01:18:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	121694	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	176388	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	165726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	82759	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	73882	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
35) Dibromofluoromethane	5.51	113	58836	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	8.71	98	221527	55.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.08%	
62) 4-Bromofluorobenzene	11.14	95	78325	54.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.24%	
Target Compounds						
16) Acetone	2.44	43	5335	7.08	ug/l	96
20) Methylene Chloride	2.85	84	4633	3.74	ug/l	88
43) Isopropyl Acetate	6.47	43	16739	5.17	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	4184	1.03	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005978.D
Acq On : 14 Nov 2018 17:02
Operator : JC/MD
Sample : PB114787ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114787ZHE#01

Quant Time: Nov 15 01:18:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

