

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005985.D
 Acq On : 14 Nov 2018 20:08
 Operator : JC/MD
 Sample : J5921-08DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-UST-MIDDLEDL

Quant Time: Nov 15 12:10:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	125026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	181834	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	161875	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	83023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	76553	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
35) Dibromofluoromethane	5.49	113	60975	49.30	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	8.71	98	216369	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
62) 4-Bromofluorobenzene	11.14	95	77797	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
Target Compounds						
31) Cyclohexane	5.57	56	10845	5.42	ug/l	88
39) Methylcyclohexane	7.46	83	6910	3.89	ug/l	90
40) Benzene	6.14	78	6633	1.30	ug/l	97
68) m/p-Xylenes	10.36	106	16243	8.19	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	175784	39.01	ug/l	99
95) Naphthalene	13.83	128	56644	10.44	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005985.D
Acq On : 14 Nov 2018 20:08
Operator : JC/MD
Sample : J5921-08DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-UST-MIDDLEDL

Quant Time: Nov 15 12:10:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

