

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005986.D
 Acq On : 14 Nov 2018 20:35
 Operator : JC/MD
 Sample : J5921-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-TANK2-N48963

Quant Time: Nov 15 03:14:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	127113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	191328	50.00	ug/l	0.01
63) Chlorobenzene-d5	10.12	117	162359	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	79877	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	74108	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	5.51	113	59149	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
50) Toluene-d8	8.71	98	215622	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.14	95	74537	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
16) Acetone	2.44	43	18371	23.35	ug/l	99
25) 2-Butanone	4.70	43	5010	4.05	ug/l	100
52) Toluene	8.79	92	2567	0.85	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005986.D
Acq On : 14 Nov 2018 20:35
Operator : JC/MD
Sample : J5921-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-TANK2-N48963

Quant Time: Nov 15 03:14:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

