

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005998.D
 Acq On : 15 Nov 2018 02:47
 Operator : JC/MD
 Sample : J5929-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 207-WEST-TRENCH

Quant Time: Nov 15 06:42:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	122139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	191391	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	152241	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	70060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	69468	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
35) Dibromofluoromethane	5.50	113	59968	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
50) Toluene-d8	8.71	98	202581	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
62) 4-Bromofluorobenzene	11.14	95	68122	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
Target Compounds						
16) Acetone	2.45	43	6160	8.15	ug/l	97
20) Methylene Chloride	2.86	84	7902	6.36	ug/l	98
43) Isopropyl Acetate	6.46	43	8379	2.38	ug/l	97
95) Naphthalene	13.83	128	3457	1.85	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005998.D
Acq On : 15 Nov 2018 02:47
Operator : JC/MD
Sample : J5929-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
207-WEST-TRENCH

Quant Time: Nov 15 06:42:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

