

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX005999.D
 Acq On : 15 Nov 2018 03:13
 Operator : JC/MD
 Sample : J5930-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 32727-45914-21716

Quant Time: Nov 15 06:44:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	127009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	182136	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	158904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	77602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71274	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
35) Dibromofluoromethane	5.51	113	63837	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
50) Toluene-d8	8.71	98	211318	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.13	95	73625	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
Target Compounds						
16) Acetone	2.45	43	6727	8.56	ug/l	92
20) Methylene Chloride	2.85	84	9295	7.20	ug/l	92
43) Isopropyl Acetate	6.46	43	9803	2.93	ug/l	99
48) Methyl methacrylate	7.77	41	44112	29.36	ug/l	99
95) Naphthalene	13.83	128	65269	12.59	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX005999.D
Acq On : 15 Nov 2018 03:13
Operator : JC/MD
Sample : J5930-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
32727-45914-21716

Quant Time: Nov 15 06:44:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

