

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX006005.D
 Acq On : 15 Nov 2018 05:53
 Operator : JC/MD
 Sample : J5860-23ME 20X
 Misc : 5.18g/5ml/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EP2-SB-110(24-28)ME

Quant Time: Nov 15 07:06:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	137995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	210839	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	181763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	81160	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	74828	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
35) Dibromofluoromethane	5.50	113	66037	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
50) Toluene-d8	8.71	98	216095	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
62) 4-Bromofluorobenzene	11.14	95	74259	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
Target Compounds						
30) Chloroform	5.21	83	3382	1.38	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX006005.D
Acq On : 15 Nov 2018 05:53
Operator : JC/MD
Sample : J5860-23ME 20X
Misc : 5.18g/5ml/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EP2-SB-110(24-28)ME

Quant Time: Nov 15 07:06:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

