

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX006011.D
 Acq On : 15 Nov 2018 08:32
 Operator : JC/MD
 Sample : J5918-07
 Misc : 5.00g/10ml/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FDSB-11(11-13)

Quant Time: Nov 15 10:05:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	126851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	193993	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	175714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	94185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	78438	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
35) Dibromofluoromethane	5.51	113	60710	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
50) Toluene-d8	8.71	98	233637	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
62) 4-Bromofluorobenzene	11.14	95	88409	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
Target Compounds						
39) Methylcyclohexane	7.45	83	1595	0.84	ug/l	# 84
73) Isopropylbenzene	11.02	105	5387	0.99	ug/l	98
78) n-propylbenzene	11.36	91	14725	2.29	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	5402	0.94	ug/l	96
85) sec-Butylbenzene	11.95	105	16207	2.83	ug/l	82
89) n-Butylbenzene	12.39	91	19681	4.19	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX006011.D
Acq On : 15 Nov 2018 08:32
Operator : JC/MD
Sample : J5918-07
Misc : 5.00g/10ml/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FDSB-11(11-13)

Quant Time: Nov 15 10:05:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

